



JENNIFER M. GRANHOLM  
GOVERNOR

STATE OF MICHIGAN  
DEPARTMENT OF LABOR & ECONOMIC GROWTH  
LANSING

STANLEY "SKIP" PRUSS  
DIRECTOR

November 6, 2008

Mr. Dale M. Wall  
Penner Manufacturing Inc.  
102 Grant Street  
Aurora, NE 68818

Dear Mr. Wall:

At the Construction Code Commission meeting on November 5, 2008, the Plumbing Division presented your request for approval of the Penner Patient Care, Cascade Bath Systems, Premier Model 360020-1P.

The Construction Code Commission, upon recommendation of the State Plumbing Board, approved the use of the Penner Patient Care, Cascade Bath Systems, Premier Model 360020-1P with the following conditions:

1. All requirements of the Michigan Plumbing and Electrical codes shall be applicable.
2. Installations shall be in accordance with the manufacturer's specifications.
3. Requests from hospital or care institutions must be approved by the Health Department having jurisdiction.
4. Both hot and cold water supplies shall be protected by and approved for reduced pressure principle backflow preventer.
5. The unit shall discharge into an approved waste receptor.

Should any change be made in the design or designation of these products, the modified products must be submitted to this office and a new acceptance obtained.

*Providing for Michigan's Safety in the Built Environment*

BUREAU OF CONSTRUCTION CODES  
P.O. BOX 30254 • LANSING, MICHIGAN 48909  
Telephone (517) 241-9330 • Fax (517) 373-8547  
[www.michigan.gov/dleg](http://www.michigan.gov/dleg)

DLEG is an equal opportunity employer/program.


Auxiliary aids, services and other reasonable accommodations are available upon request to individuals with disabilities.

Mr. Dale M. Wall  
Page 2  
November 6, 2008

In accordance with Section 21 of 1972 PA 230, neither this letter nor the Certificate of Acceptability may be used for advertising or promotional purposes.

A Certificate of Acceptability is enclosed.

Sincerely,

A handwritten signature in black ink, appearing to read "Irvin J. Poke". The signature is fluid and cursive, with a long, sweeping underline that extends to the right.

Irvin J. Poke, AIA  
Director

RGK/rgk

Enclosure

cc: Duane Branch, Chairperson, State Plumbing Board

**CERTIFICATE OF ACCEPTABILITY**

**NO. 1593 PA**

Issued by the Michigan State Construction Code Commission in accordance with authority granted under Section 21 of 1972 PA 230, being Section 125.1521 of the Michigan Compiled Laws on the recommendation of the State Plumbing Board.

**MANUFACTURER:** Penner Manufacturing Inc.  
102 Grant Street  
Aurora, NE 68818


**PRODUCT:** Penner Patient Care, Cascade Bath Systems, Premier, Model 360020-1P

**CONDITIONS OF USE AND INSTALLATION:**

1. All requirements of the Michigan Plumbing and Electrical codes shall be applicable.
2. Installations shall be in accordance with the manufacturer's specifications.
3. Requests from hospital or care institutions must be approved by the Health Department having jurisdiction.
4. Both hot and cold water supplies shall be protected by and approved for reduced pressure principle backflow preventer.
5. The unit shall discharge into an approved waste receptor.

**THIS CERTIFICATE SHALL NOT BE USED FOR ADVERTISING PURPOSES.**

By Direction of the Construction Code Commission



Irvin J. Poke, AIA, Director, Bureau of Construction Codes

  
Robert G. Konyndyk, Chief, Plumbing Division

Effective Date: November 5, 2008