



CALIBRATION LABORATORIES

NVLAP LAB CODE 200408-0

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005**

<p><b>Michigan Dept. of Agriculture &amp; Rural Dev, E.C. Heffron Metrology Laboratory</b> 940 Venture Lane Williamston, MI 48895 Mr. Nicholas Santini Phone: 517-655-8202 Fax: 517-655-8303 E-mail: santinin@michigan.gov URL: www.michigan.gov/wminfo</p>	<p><b>Fields of Calibration</b> Mechanical</p> <p>This laboratory is compliant to ANSI/NCSL Z540-1-1994; Part 1. (NVLAP Code: 20/A01)</p>
---	---

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)<sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3</sup>	Remarks
<b>MECHANICAL</b>			
<b>FORCE (20/M06)</b>			
	20 000 lbf 15 000 lbf 10 000 lbf 5 000 lbf	32 lbf 25 lbf 27 lbf 27 lbf	Wheel Load Weighers
<b>MASS (20/M08)</b>			
Metric	20 kg 10 kg 5 kg 3 kg 2 kg 1 kg 500 g 300 g 200 g 100 g 50 g 30 g 20 g 10 g 5 g 3 g 2 g	22 mg 1.4 mg 0.72 mg 0.48 mg 0.32 mg 80 µg 43 µg 28 µg 22 µg 20 µg 10 µg 6.6 µg 4.7 µg 3.8 µg 2.1 µg 1.5 µg 1.2 µg	Echelon I

2016-08-23 through 2017-09-30  
Effective dates

For the National Voluntary Laboratory Accreditation Program



**National Voluntary  
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200408-0

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)** <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3</sup>	Remarks
	1 g	1.2 µg	Echelon II
	500 mg	0.87 µg	
	300 mg	0.71 µg	
	200 mg	0.67 µg	
	100 mg	0.77 µg	
	50 mg	0.40 µg	
	30 mg	0.27 µg	
	20 mg	0.20 µg	
	10 mg	0.18 µg	
	5 mg	0.17 µg	
	3 mg	0.18 µg	
	2 mg	0.19 µg	
	1 mg	0.22 µg	
	20 kg	30 mg	
	10 kg	6.8 mg	
	5 kg	1.3 mg	
	3 kg	1.0 mg	
	2 kg	1.3 mg	
	1 kg	0.18 mg	
	500 g	0.32 mg	
	300 g	0.19 mg	
	200 g	0.13 mg	
	100 g	70 µg	
	50 g	30 µg	
	30 g	22 µg	
	20 g	17 µg	
	10 g	11 µg	
	5 g	4.7 µg	
	3 g	3.2 µg	
	2 g	2.5 µg	
	1 g	1.9 µg	
	500 mg	1.9 µg	
	300 mg	1.6 µg	
	200 mg	1.1 µg	
	100 mg	1.1 µg	

2016-08-23 through 2017-09-30

Effective dates

For the National Voluntary Laboratory Accreditation Program



**National Voluntary  
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200408-0

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)** <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3</sup>	Remarks
Avoirdupois	50 mg	0.81 µg	Echelon II
	30 mg	0.86 µg	
	20 mg	0.62 µg	
	10 mg	0.79 µg	
	5 mg	0.48 µg	
	3 mg	0.42 µg	
	2 mg	0.45 µg	
	1 mg	0.44 µg	
	50 lb	30 mg	
	25 lb	20 mg	
	20 lb	6.8 mg	
	10 lb	1.3 mg	
	5 lb	0.92 mg	
	3 lb	1.3 mg	
	2 lb	0.20 mg	
	1 lb	0.32 mg	
	0.5 lb	0.13 mg	
	0.3 lb	0.13 mg	
	0.2 lb	71 µg	
	0.1 lb	41 µg	
	0.05 lb	18 µg	
	0.03 lb	12 µg	
	0.02 lb	12 µg	
	0.01 lb	5.3 µg	
	0.005 lb	3.1 µg	
	0.003 lb	2.7 µg	
	0.002 lb	2.4 µg	
	0.001 lb	2.1 µg	
	8 oz	0.13 mg	
	4 oz	0.13 mg	
2 oz	41 µg		
1 oz	23 µg		
1/2 oz	12 µg		
1/4 oz	12 µg		
1/8 oz	3.9 µg		

2016-08-23 through 2017-09-30

Effective dates

For the National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200408-0

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)<sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3</sup>	Remarks
Metric	1/16 oz	3.1 µg	Echelon III
	1/32 oz	2.4 µg	
	500 kg	2.0 g	
	250 kg	1.8 g	
	100 kg	1.8 g	
	50 kg	0.25 g	
	25 kg	89 mg	
	20 kg	0.11 g	
	10 kg	0.10 g	
	5 kg	5.7 mg	
	3 kg	4.7 mg	
	2 kg	4.9 mg	
	1 kg	4.5 mg	
	500 g	4.4 mg	
	300 g	4.5 mg	
	200 g	0.23 mg	
	100 g	0.13 mg	
	50 g	0.10 mg	
	30 g	75 µg	
	20 g	73 µg	
	10 g	73 µg	
	5 g	72 µg	
	3 g	72 µg	
	2 g	70 µg	
	1 g	75 µg	
	500 mg	9.1 µg	
	300 mg	9.2 µg	
	200 mg	9.3 µg	
	100 mg	8.8 µg	
	50 mg	8.5 µg	
30 mg	8.6 µg		
20 mg	8.6 µg		
10 mg	8.6 µg		
5 mg	8.5 µg		
3 mg	8.5 µg		

2016-08-23 through 2017-09-30

Effective dates

For the National Voluntary Laboratory Accreditation Program



**National Voluntary  
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200408-0

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)** <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3</sup>	Remarks
Avoirdupois	2 mg	8.6 µg	Echelon III
	1 mg	8.4 µg	
	5000 lb	0.20 lb	
	1000 lb	4.5 µlb	
	500 lb	4.0 µlb	
	100 lb	0.27 g	
	50 lb	0.10 g	
	25 lb	83 mg	
	20 lb	80 mg	
	10 lb	5.5 mg	
	5 lb	4.5 mg	
	3 lb	5.1 mg	
	2 lb	4.3 mg	
	1 lb	4.5 mg	
	0.5 lb	4.0 mg	
	0.3 lb	0.20 mg	
	0.2 lb	0.13 mg	
	0.1 lb	0.10 mg	
	0.05 lb	84 µg	
	0.03 lb	80 µg	
	0.02 lb	78 µg	
	0.01 lb	77 µg	
	0.005 lb	88 µg	
	0.003 lb	84 µg	
	0.002 lb	81 µg	
	0.001 lb	81 µg	
	8 oz	4.0 mg	
	4 oz	0.21 mg	
2 oz	0.11 mg		
1 oz	87 µg		
1/2 oz	78 µg		
1/4 oz	80 µg		
1/8 oz	77 µg		
1/16 oz	88 µg		

2016-08-23 through 2017-09-30

Effective dates

For the National Voluntary Laboratory Accreditation Program



**National Voluntary  
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200408-0

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)** <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty <sup>Note 3</sup>	Remarks
<b>VOLUME AND DENSITY (20/M12)</b>			
	2000 gal	92 in <sup>3</sup>	Volume Transfer
	1500 gal	67 in <sup>3</sup>	
	1000 gal	48 in <sup>3</sup>	
	750 gal	27 in <sup>3</sup>	
	500 gal	19 in <sup>3</sup>	
	100 gal	2.9 in <sup>3</sup>	
	50 gal	2.1 in <sup>3</sup>	
	5 gal	0.27 in <sup>3</sup>	
	25 gal	0.42 in <sup>3</sup>	Volume Gravimetric
	15 gal	0.39 in <sup>3</sup>	
	5 gal	0.11 in <sup>3</sup>	
	1 gal	0.042 in <sup>3</sup>	
	1/2 gal	0.041 in <sup>3</sup>	
	1 qt	0.020 in <sup>3</sup>	
	1 pt	0.015 in <sup>3</sup>	
	1/2 pt	0.0048 in <sup>3</sup>	
	30 gal	0.79 in <sup>3</sup>	Small Volume Prover
	20 gal	0.35 in <sup>3</sup>	
	15 gal	0.39 in <sup>3</sup>	
	5 gal	0.011 in <sup>3</sup>	
	2 liter	0.65 mL	Volume Gravimetric
	1 liter	0.35 mL	
	100 mL	0.074 mL	
<b>END</b>			

2016-08-23 through 2017-09-30

Effective dates

For the National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200408-0

Notes

**Note 1:** A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

**Note 2:** Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

**Note 3:** The uncertainty associated with a measurement in a CMC is an expanded uncertainty with a level of confidence of approximately 95 %, typically using a coverage factor of  $k = 2$ . However, laboratories may report a coverage factor different than  $k = 2$  to achieve the 95 % level of confidence. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

**Note 3a:** The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

**Note 3b:** As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

**Note 3c:** As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.1.h. of NIST Handbook 150, Procedures and General Requirements.

**Note 4:** Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

**Note 5:** Values listed with percent (%) are percent of reading or generated value unless otherwise noted.

**Note 6:** NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

2016-08-23 through 2017-09-30

Effective dates

For the National Voluntary Laboratory Accreditation Program