



Medical Director: Responsibilities and Duties

PROCEDURE

MEDICAL DIRECTOR'S QUALITY ASSURANCE REVIEW

OBJECTIVE

All mammography facilities shall designate a licensed physician to provide medical direction for the delivery of mammography services. The objective of this form is to provide the designated physician ("medical director") with guidance and information to help the medical director perform and document the required responsibilities.

The medical director is responsible for the clinical aspects of the x-ray examinations and other procedures related to mammography. Also, the medical director is responsible for conducting an on-site visit to each mammography station within the facility at least monthly for the purpose of providing professional feedback regarding clinical image quality and quality assurance procedures, for review of quality control documentation, and for ensuring safe operating procedures are used in the delivery of mammography services (Section 13523(2)(d)(iii) of the Public Health Code of the State of Michigan).

In conjunction with this form, we highly recommend that the medical director be familiar with all sections of the American College of Radiology's Mammography Quality Control Manual.

The medical director shall provide **monthly** professional feedback regarding clinical image quality, safe operating procedures, and quality control documentation.

FREQUENCY

Not all quality control tests are performed on a monthly basis. Some quality control tests are performed quarterly, semiannually, or annually. The medical director's review of each quality control test should be done the month immediately following when the test was performed.

PROCEDURE STEPS

1. If the medical director practices primarily outside of the facility, the physician must visit the mammography facility at least once a month for the purpose of providing professional feedback regarding clinical image quality and quality assurance procedures, for review of quality control documentation, and for ensuring safe operating procedures are used in the delivery of mammography services.



Medical Director: Responsibilities and Duties

PROCEDURE STEPS (Continued)

2. The medical director shall review the technologist's quality control tests, ensuring that:
 - The test was performed at the required frequency.
 - The test was performed using the written procedure in the facility's quality assurance manual.
 - The results are properly documented or charted.
 - When the test did not fall within the limits of acceptability, established in the facility's quality assurance procedures manual, then the protocol for corrective action, which is also written in the facility's quality assurance procedures manual was followed.

The medical director needs to be familiar enough with the facility's quality assurance procedures manual to review and critique the results of any of the quality control tests. The review is to ensure tests were done properly, in a timely manner, and that test results were either within the control limits or that appropriate corrective action was taken.

3. The medical director shall review the medical physicist's report, ensuring that any item found that does not meet the requirements of Michigan's mammography rules has been corrected and documented.
4. The medical director shall also ensure that:
 - Clinical aspects of the x-ray examinations and other procedures related to mammography are appropriate
 - The clinical image quality of mammograms is appropriate
 - Safe operating procedures are used in the delivery of mammography services

The medical director should use the following checklist, or one designed by the facility, on a monthly basis to document the on-site visit and that necessary reviews have been made.



Medical Director: Responsibilities and Duties

Mammography Quality Assurance Check List

Medical Director's Monthly On-Site Checks

Year												
Month	Jan	Feb	Mar	April	May	June	July	Aug	Sept	Oct	Nov	Dec
Technologist's QC Tests												
Physicist's report (annually)												
Clinical image quality is appropriate												
Safe operating procedures are used												
Clinical aspects of the x-ray exams and other procedures related to mammography are appropriate												

The appropriate box should be dated and initialed by the medical director to certify when that particular review is completed during the on-site visit.

This is a monthly duty