

MICHIGAN DEPARTMENT OF COMMUNITY HEALTH
CERTIFICATE OF NEED (CON) REVIEW STANDARDS FOR
COMPUTED TOMOGRAPHY (CT) SCANNER SERVICES

(By authority conferred on the CON Commission by Section 22215 of Act No. 368 of the Public Acts of 1978, as amended, and sections 7 and 8 of Act No. 306 of the Public Acts of 1969, as amended, being sections 333.22215, 24.207 and 24.208 of the Michigan Compiled Laws.)

Section 1. Applicability

Sec. 1. These standards are requirements for the approval of the initiation, expansion, replacement, ~~relocation~~, or acquisition of CT services and the delivery of services under Part 222 of the Code. Pursuant to Part 222 of the Code, CT is a covered clinical service. The Department shall use these standards in applying Section 22225(1) of the Code, being Section 333.22225(1) of the Michigan Compiled Laws and Section 22225(2)(c) of the Code, being Section 333.22225(2)(c) of the Michigan Compiled Laws.

Section 2. Definitions

Sec. 2. (1) For purposes of these standards:

(a) "Acquisition of an existing CT scanner service" means obtaining possession or control of an existing fixed or mobile CT scanner service or existing CT scanner(s) by contract, ownership, or other comparable arrangement. For proposed projects involving mobile CT scanners, this applies to the central service coordinator and/or host facility.

(b) "Billable procedure" means a CT procedure billed as a single unit under procedure codes in effect on December 31, 2010, and performed in Michigan.

(c) "Body scans" include all spinal CT scans and any CT scan of an anatomical site below and including the neck.

(d) "Central service coordinator" means the organizational unit which has operational responsibility for a mobile CT scanner and which is a legal entity authorized to do business in the state of Michigan.

(e) "Certificate of Need Commission" or "Commission" means the Commission created pursuant to Section 22211 of the Code, being Section 333.22211 of the Michigan Compiled Laws.

(f) "Code" means Act No. 368 of the Public Acts of 1978, as amended, being Section 333.1101 et seq. of the Michigan Compiled Laws.

(g) "Computed tomography" or "CT" means the use of radiographic and computer techniques to produce cross-sectional images of the head or body.

(h) "CT-ANGIO HYBRID UNIT" MEANS AN INTEGRATED SYSTEM COMPRISING BOTH CT AND ANGIOGRAPHY EQUIPMENT SITED IN THE SAME ROOM THAT IS DESIGNED SPECIFICALLY FOR INTERVENTIONAL RADIOLOGY OR CARDIAC PROCEDURES. THE CT UNIT IS A GUIDANCE MECHANISM AND IS INTENDED TO BE USED AS AN ADJUNCT TO THE PROCEDURE. THE CT UNIT SHALL NOT BE USED FOR DIAGNOSTIC STUDIES UNLESS THE PATIENT IS CURRENTLY UNDERGOING A CT-ANGIO HYBRID PROCEDURE AND IS IN NEED OF A SECONDARY DIAGNOSTIC STUDY.

Formatted: Highlight

Formatted: Highlight

Formatted: Highlight

(i) "CT equivalents" means the resulting number of units produced when the number of billable procedures for each category is multiplied by its respective conversion factor tabled in Section 21.

(ii) "CT scanner" means x-ray CT scanning systems capable of performing CT scans of the head, other body parts, or full body patient procedures including Positron Emission Tomography (PET)/CT scanner hybrids if used for CT only procedures. The term does not include emission-computed tomographic systems utilizing internally administered single-photon gamma ray emitters, positron annihilation CT systems, magnetic resonance, ultrasound computed tomographic systems, CT simulators used solely for treatment planning purposes in conjunction with an MRT unit, and non-diagnostic, intra-operative guidance tomographic units.

(k) "CT scanner services" means the CON-approved utilization of a CT scanner(s) at one site in the case of a fixed CT scanner service or at each host site in the case of a mobile CT scanner service.

55 | (k) "Dedicated pediatric CT" means a fixed CT scanner on which at least 70% of the CT procedures
56 | are performed on patients under 18 years of age.

57 | (lm) "Dental CT examinations" means use of a CT scanner specially designed to generate CT images
58 | to facilitate dental procedures.

59 | (mn) "Dental procedures" means dental implants, wisdom teeth surgical procedures, mandibular or
60 | maxillary surgical procedures, or temporal mandibular joint evaluations.

61 | (po) "Department" means the Michigan Department of Community Health (MDCH).

62 | (ep) "Emergency room" means a designated area physically part of a licensed hospital and recognized
63 | by the Department as having met the staffing and equipment requirements for the treatment of emergency
64 | patients.

65 | (pq) "Excess CT Equivalents" means the number of CT equivalents performed by an existing CT
66 | scanner service in excess of 10,000 per fixed CT scanner and 4,500 per mobile CT scanner or either an
67 | existing fixed or mobile CT scanner service, the number of CT scanners used to compute excess CT
68 | equivalents shall include both existing and approved but not yet operational CT scanners. In the case of a
69 | CT scanner service that operates or has a valid CON to operate that has more than one fixed CT scanner
70 | at the same site, the term means number of CT equivalents in excess of 10,000 multiplied by the number
71 | of fixed CT scanners at the same site. For example, if a CT scanner service operates, or has a valid CON
72 | to operate, two fixed CT scanners at the same site, the excess CT equivalents is the number that is in
73 | excess of 20,000 (10,000 x 2) CT equivalents. In the case of an existing mobile CT scanner service, the
74 | term means the sum of all CT equivalents performed by the same mobile CT scanner service at all of the
75 | host sites combined that is in excess of 4,500. For example, if a mobile CT scanner service serves five
76 | host sites with 1 mobile CT scanner, the term means the sum of CT equivalents for all five host sites
77 | combined that is in excess of 4,500 CT equivalents.

78 | (qr) "Existing CT scanner service" means the utilization of a CON-approved and operational CT
79 | scanner(s) at one site in the case of a fixed CT scanner service or at each host site in the case of a
80 | mobile CT scanner service.

81 | (rs) "Existing CT scanner" means a CON-approved and operational CT scanner used to provide CT
82 | scanner services.

83 | (st) "Existing mobile CT scanner service" means a CON-approved and operational CT scanner and
84 | transporting equipment operated by a central service coordinator serving two or more host sites.

85 | (t) "Expand an existing CT scanner service" means the addition of one or more CT scanners at an
86 | existing CT scanner service.

87 | (u) "Head scans" include head or brain CT scans; including the maxillofacial area; the orbit, sella, or
88 | posterior fossa; or the outer, middle, or inner ear; or any other CT scan occurring above the neck.

89 | (vu) "Health Service Area" or "HSA" means the groups of counties listed in [Section 24 APPENDIX A](#).

90 | (wv) "HIPAA" means the Health Insurance Portability and Accountability Act of 1996.

91 | (xw) "Hospital-based portable CT scanner or portable CT scanner" means a CT scanner capable of
92 | being transported into patient care areas (i.e., ICU rooms, operating rooms, etc.) to provide high-quality
93 | imaging of critically ill patients.

94 | (yx) "Host site" means the site at which a mobile CT scanner is authorized to provide CT scanner
95 | services.

96 | (zy) "Initiate a CT scanner service" means to begin operation of a CT scanner, whether fixed or
97 | mobile, at a site that does not perform CT scans as of the date an application is submitted to the
98 | Department. The term does not include the acquisition or ~~relocation~~ [REPLACEMENT](#) of an existing CT
99 | scanner service [TO A DIFFERENT SITE](#) or the renewal of a lease.

100 | (az) "Medicaid" means title XIX of the social security act, chapter 531, 49 Stat. 620, 1396 ~~f-6 and 1396f-~~
101 | ~~8 to 1396v~~ [1396w-5](#).

102 | ~~(bb) "Metropolitan statistical area county" means a county located in a metropolitan statistical area as~~
103 | ~~that term is defined under the "standards for defining metropolitan and micropolitan statistical areas" by~~
104 | ~~the statistical policy office of the office of information and regulatory affairs of the United States office of~~
105 | ~~management and budget, 65 F.R. p. 82238 (December 27, 2000) and as shown in Appendix A.~~
106 | ~~(cc) "Micropolitan statistical area county" means a county located in a micropolitan statistical area as~~
107 | ~~that term is defined under the "standards for defining metropolitan and micropolitan statistical areas" by~~

~~the statistical policy office of the office of information and regulatory affairs of the United States office of management and budget, 65 F.R. p. 82238 (December 27, 2000) and as shown in Appendix A.~~
(~~ddaa~~) "Mobile CT scanner service" means a CT scanner and transporting equipment operated by a central service coordinator and which must serve two or more host facilities.
(~~eebb~~) "Mobile CT scanner network" means the route (all host facilities) the mobile CT scanner is authorized to serve.
(#cc) "Pediatric patient" means any patient less than 18 years of age.
~~(gg) "Relocate a fixed CT scanner" means a change in the location of a fixed CT scanner from the existing site to a different site within the relocation zone.~~
(hh) "Relocate an existing CT scanner service" means a change in the geographic location of an existing fixed CT scanner service from an existing site to a different site.
(ii) "Relocation zone," means a site that is within a 10-mile radius of a site at which an existing fixed CT scanner service is located if an existing fixed CT scanner service is located in a metropolitan statistical area county, or a 20-mile radius if an existing fixed CT scanner service is located in a rural or micropolitan statistical area county.
(jdd) "Replace an existing CT scanner" means an equipment change of an existing CT scanner, that requires a change in the radiation safety certificate, proposed by an applicant which results in that applicant operating the same number of CT scanners before and after project completion, at the same geographic location. THE TERM ALSO INCLUDES RELOCATING an existing CT scanner OR CT SCANNER service from an existing site to a different site.
~~(kk) "Rural county" means a county not located in a metropolitan statistical area or micropolitan statistical areas as those terms are defined under the "standards for defining metropolitan and micropolitan statistical areas" by the statistical policy office of the office of information regulatory affairs of the United States office of management and budget, 65 F.R. p. 82238 (December 27, 2000) and as shown in Appendix A.~~
(#ee) "Sedated patient" means a patient that meets all of the following:
(i) Patient undergoes procedural sedation and whose level of consciousness is either moderate sedation or a higher level of sedation, as defined by the American Association of Anesthesiologists, the American Academy of Pediatrics, the Joint Commission on the Accreditation of Health Care Organizations, or an equivalent definition.
(ii) Who requires observation by personnel, other than technical employees routinely assigned to the CT unit, who are trained in cardiopulmonary resuscitation (CPR) and pediatric advanced life support (PALS).
(~~mmff~~) "Special needs patient" means a non-sedated patient, either pediatric or adult, with any of the following conditions: down syndrome, autism, attention deficit hyperactivity disorder (ADHD), developmental delay, malformation syndromes, hunter's syndrome, multi-system disorders, psychiatric disorders, and other conditions that make the patient unable to comply with the positional requirements of the exam.

(2) Terms defined in the Code have the same meanings when used in these standards.

Section 3. Requirements for approval for applicants proposing to initiate a CT scanner service other than a dental CT scanner service or hospital-based portable CT scanner service

Sec. 3. An applicant proposing to initiate a CT scanner service shall demonstrate each of the following, as applicable:

(1) A hospital proposing to initiate its first fixed CT scanner service shall demonstrate each of the following:

- (a) The proposed site is a hospital licensed under Part 215 of the Code.
- (b) The hospital operates an emergency room that provides 24-hour emergency care services as authorized by the local medical control authority to receive ambulance runs.

(2) An applicant, other than an applicant meeting all of the applicable requirements of subsection (1), proposing to initiate a fixed CT scanner service shall project an operating level of at least 7,500 CT equivalents per year for the second 12-month period after beginning operation of the CT scanner.

(3) An applicant proposing to initiate a mobile CT scanner service shall project an operating level of at least 3,500 CT equivalents per year for the second 12-month period after beginning operation of the CT scanner.

(4) AN APPLICANT PROPOSING TO INITIATE CT SCANNER SERVICES AS AN EXISTING HOST SITE ON A DIFFERENT MOBILE CT SCANNER SERVICE SHALL DEMONSTRATE THE FOLLOWING:

(a) THE APPLICANT PROVIDES A PROPOSED ROUTE SCHEDULE.

(b) THE APPLICANT PROVIDES A DRAFT CONTRACT FOR SERVICES BETWEEN THE PROPOSED HOST SITE AND CENTRAL SERVICE COORDINATOR.

Comment [A1]: Response to #4 CT Workgroup charge and evaluation section.

Section 4. Requirements for approval for applicants proposing to initiate a dental CT scanner service

Sec. 4. An applicant proposing to initiate A FIXED OR MOBILE dental CT scanner service shall demonstrate each of the following, as applicable:

(1) An applicant is proposing a DENTAL ~~fixed~~ CT scanner service for the sole purpose of performing dental CT examinations.

Comment [A2]: Response to #1 CT Workgroup Charge

(2) The CT scanner generates a peak power of 5 kilowatts or less as certified by the manufacturer.

(3) An applicant proposing to initiate a dental CT scanner service, other than an applicant that is proposing a dental CT scanner service in HSA 8, shall project an operating level of at least 200 dental CT examinations per year for the second 12-month period after beginning operation of the dental CT scanner.

(4) The applicant has demonstrated to the satisfaction of the Department that the person(s) (e.g., technician, dentist) operating the dental CT scanner has been appropriately trained and/or certified by one of the following groups, as recognized by the Department: a dental radiology program in a certified dental school, an appropriate professional society, or a dental continuing education program accredited by the American Dental Association.

(5) The applicant has demonstrated to the satisfaction of the Department that the dental CT examinations generated by the proposed dental CT scanner will be interpreted by a licensed dentist(s) trained and/or certified by one of the following groups, as recognized by the Department: a dental radiology program in a certified dental school, an appropriate professional society, or a dental continuing education program accredited by the American Dental Association.

Section 5. Requirements for approval for applicants proposing to expand an existing CT scanner service other than a dental CT scanner service or hospital-based portable CT scanner service

Sec. 5. (1) An applicant proposing to expand an existing fixed CT scanner service shall demonstrate that all of the applicant's fixed CT scanners, excluding CT scanners approved pursuant to sections 13 and 17, have performed an average of at least 10,000 CT equivalents per fixed CT scanner for the most recent continuous 12-month period preceding the applicant's request. In computing this average, the Department will divide the total number of CT equivalents performed by the applicant's total number of fixed CT scanners, including both operational and approved but not operational fixed CT scanners.

(2) An applicant proposing to expand an existing fixed CT scanner service approved pursuant to Section 17 shall demonstrate that all of the applicant's dedicated pediatric CT scanners have performed an average of at least 3,000 CT equivalents per dedicated pediatric CT scanner for the most recent

continuous 12-month period preceding the applicant's request. In computing this average, the Department will divide the total number of CT equivalents performed by the applicant's total number of dedicated pediatric CT scanners, including both operational and approved but not operational dedicated pediatric CT scanners.

(3) If an applicant proposes to expand an existing mobile CT scanner service, the applicant shall demonstrate that all of the applicant's mobile CT scanners have performed an average of at least 5,500 CT equivalents per mobile CT scanner for the most recent continuous 12-month period preceding the applicant's request. In computing this average, the Department will divide the total number of CT equivalents performed by the applicant's total number of mobile CT scanners, including both operational and approved but not operational mobile CT scanners.

Section 6. Requirements for approval for applicants proposing to expand an existing dental CT scanner service

Sec. 6. An applicant proposing to expand an existing fixed OR MOBILE dental CT scanner service shall demonstrate that all of the applicant's dental CT scanners have performed an average of at least 300 dental CT examinations per fixed OR MOBILE dental CT scanner for the most recent continuous 12-month period preceding the applicant's request. In computing this average, the Department will divide the total number of dental CT examinations performed by the applicant's total number of fixed OR MOBILE dental CT scanners, including both operational and approved but not operational fixed OR MOBILE dental CT scanners.

Comment [A3]: Response to #1 CT Workgroup Charge

Section 7. Requirements for approval for applicants proposing to replace an existing CT scanner other than a dental CT scanner or hospital-based portable CT scanner

Sec. 7. An applicant proposing to replace an existing CT scanner, EXCEPT FOR AN APPLICANT APPROVED UNDER SECTION 3(1), shall demonstrate each of the following, as applicable:

(1) An applicant, other than an applicant meeting all of the applicable requirements of subsection (a), (b) or (c) below, proposing to replace an existing fixed CT scanner shall demonstrate that the fixed CT scanner(s) performed at least an average of 7,500 CT equivalents per fixed CT scanner in the most recent 12-month period for which the Department has verifiable data.

(a) A hospital proposing to replace an existing CT scanner which is the only fixed CT scanner operated at that site by the hospital shall demonstrate each of the following:

(i) The proposed site is a hospital licensed under Part 215 of the Code.

(ii) The hospital operates an emergency room that provides 24-hour emergency care services as authorized by the local medical control authority to receive ambulance runs.

(iii) The replacement CT scanner will be located at the same site as the CT scanner to be replaced.

(b) An applicant proposing to replace an existing fixed CT scanner shall be exempt once from the volume requirements if the existing CT scanner demonstrates that it meets all of the following:

(i) The existing CT scanner has performed at least 5,000 CT equivalents in the most recent 12-month period for which the Department has verifiable data.

(ii) The existing CT scanner is fully depreciated according to generally accepted accounting principles.

(iii) The existing CT scanner has at one time met its minimum volume requirements.

(c) An applicant proposing to replace an existing fixed CT scanner on an academic medical center campus, at the same site, shall be exempt once, as of May 5, 2008, from the minimum volume requirements for replacement if the existing CT scanner is fully depreciated according to generally accepted accounting principles.

(d) An applicant proposing to replace an existing fixed OR MOBILE CT scanner having a configuration of less than 16 multi-detector rows shall be exempt once, as of ~~the effective date~~ FEBRUARY 27, 2012 ~~of the standards~~, from the minimum volume requirements for replacement if it meets both of the following:

Formatted: Highlight

- (i) The proposed CT scanner to be obtained will have a configuration of sixteen (16) or more multi-detector rows, and
- (ii) The existing CT scanner is fully depreciated according to generally accepted accounting principles.
- (2) An applicant proposing to replace an existing mobile CT scanner(s) shall demonstrate that the mobile CT scanner(s) performed at least 3,500 CT equivalents if the applicant operates only one mobile CT scanner or an average of 5,500 CT equivalents for each CT scanner if the applicant operates more than one mobile CT scanner for the same mobile CT scanner network, in the most recent 12-month period for which the department has verifiable data.
- (3) An applicant proposing to replace an existing dedicated pediatric CT scanner(s) shall demonstrate that the dedicated pediatric CT scanner(s) performed at least an average of 2,500 CT equivalents per dedicated pediatric CT scanner in the most recent 12-month period for which the Department has verifiable data.
- (4) An applicant under this section shall demonstrate that the existing CT scanner(s) proposed to be replaced is fully depreciated according to generally accepted accounting principles, or, that the existing equipment clearly poses a threat to the safety of the public, or, that the proposed replacement CT scanner offers technological improvements which enhance quality of care, increase efficiency, and/or reduce operating costs and patient charges.
- (5) An applicant proposing to ~~relocate~~REPLACE an existing fixed CT scanner service TO A DIFFERENT SITE shall demonstrate that the proposed project meets all of the following:
- (a) The existing fixed CT scanner service to be ~~relocated~~REPLACED has been in operation for at least 36 months as of the date an application is submitted to the Department.
- (b) The proposed new site ~~is in the relocation zone~~ is within a 10-mile radius of a site at which an existing fixed CT scanner service is located if an existing fixed CT scanner service is located in a metropolitan statistical area county, or a 20-mile radius if an existing fixed CT scanner service is located in a rural or micropolitan statistical area county.
- (c) The requirements of sections 5 or 7, as applicable, have been met.
- (d) The CT scanner service to be relocated performed at least an average of 7,500 CT equivalents per fixed scanner in the most recent 12-month period for which the Department has verifiable data.
- (e) The applicant agrees to operate the CT scanner service in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.
- (6) An applicant proposing to ~~relocate~~REPLACE a fixed CT scanner(s) of an existing CT scanner service TO A DIFFERENT SITE shall demonstrate that the proposed project meets all of the following:
- (a) The existing CT scanner service from which the CT scanner(s) is to be relocated has been in operation for at least 36 months as of the date an application is submitted to the Department.
- (b) The proposed new site ~~is in the relocation zone~~ is within a 10-mile radius of a site at which an existing fixed CT scanner service is located if an existing fixed CT scanner service is located in a metropolitan statistical area county, or a 20-mile radius if an existing fixed CT scanner service is located in a rural or micropolitan statistical area county.
- (c) The requirements of sections 5 or 7, as applicable, have been met.
- (d) Each existing CT scanner at the service from which a scanner is to be relocated performed at least an average of 7,500 CT equivalents per fixed scanner in the most recent 12-month period for which the Department has verifiable data.
- (e) The applicant agrees to operate the CT scanner(s) at the proposed site in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.

Section 8. Requirements for approval for applicants proposing to replace an existing dental CT scanner

Sec. 8. An applicant proposing to replace an existing dental CT scanner shall demonstrate each of the following:

(1) An applicant proposing to replace an existing fixed OR MOBILE dental CT scanner shall demonstrate that the fixed OR MOBILE dental CT scanner(s) performed at least an average of 200 dental CT examinations per fixed OR MOBILE dental CT scanner in the most recent 12-month period for which the Department has verifiable data.

(2) An applicant under this section shall demonstrate that the existing dental CT scanner(s) proposed to be replaced is fully depreciated according to generally accepted accounting principles, or, that the existing equipment clearly poses a threat to the safety of the public, or that the proposed replacement dental CT scanner offers technological improvements which enhance quality of care, increase efficiency, and/or reduce operating costs and patient charges.

(3) An applicant proposing to ~~relocate~~ REPLACE an existing fixed dental CT scanner service TO A DIFFERENT SITE shall demonstrate that the proposed project meets all of the following:

(a) The existing fixed dental CT scanner service to be ~~relocated~~ REPLACED has been in operation for at least 36 month as of the date an application is submitted to the Department.

(b) FOR VOLUME PURPOSES, THE NEW SITE SHALL REMAIN ASSOCIATED WITH THE EXISTING CT SERVICE FOR A MINIMUM OF THREE YEARS.

(c) The proposed new site is ~~in the relocation ZONE~~ WITHIN A 10-MILE RADIUS OF A SITE AT WHICH AN EXISTING FIXED DENTAL CT SCANNER SERVICE IS LOCATED IF AN EXISTING FIXED DENTAL CT SCANNER SERVICE IS LOCATED IN A METROPOLITAN STATISTICAL AREA COUNTY, OR A 20-MILE RADIUS IF AN EXISTING FIXED DENTAL CT SCANNER SERVICE IS LOCATED IN A RURAL OR MICROPOLITAN STATISTICAL AREA COUNTY.

(d) The requirements of sections 6 or 8, as applicable, have been met.

(e) The dental CT scanner service to be relocated performed at least an average of 200 dental CT examinations per fixed dental CT scanner in the most recent 12-month period for which the Department has verifiable data.

(f) The applicant agrees to operate the dental CT scanner service in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.

(4) An applicant proposing to ~~relocate~~ REPLACE a fixed dental CT scanner(s) of an existing dental CT scanner service TO A DIFFERENT SITE shall demonstrate that the proposed project meets all of the following:

(a) The existing dental CT scanner service from which the dental CT scanner(s) is to be ~~relocated~~ REPLACED has been in operation for at least 36 months as of the date an application is submitted to the Department.

(b) FOR VOLUME PURPOSES, THE NEW SITE SHALL REMAIN ASSOCIATED WITH THE EXISTING CT SERVICE FOR A MINIMUM OF THREE YEARS.

(c) The proposed new site is ~~in the relocation zone~~ WITHIN A 10-MILE RADIUS OF A SITE AT WHICH AN EXISTING FIXED DENTAL CT SCANNER SERVICE IS LOCATED IF AN EXISTING FIXED DENTAL CT SCANNER SERVICE IS LOCATED IN A METROPOLITAN STATISTICAL AREA COUNTY, OR A 20-MILE RADIUS IF AN EXISTING FIXED DENTAL CT SCANNER SERVICE IS LOCATED IN A RURAL OR MICROPOLITAN STATISTICAL AREA COUNTY.

(d) The requirements of sections 6 or 8, as applicable have been met.

(e) Each existing dental CT scanner at the service from which a scanner is to be relocated performed at least an average of 200 dental CT examinations per fixed dental CT scanner in the most recent 12-month period for which the Department has verifiable data.

(f) The applicant agrees to operate the dental CT scanner(s) at the proposed site in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.

Formatted: Font: Not Bold

~~Section 9. Requirements for approval for applicants proposing to relocate an existing CT scanner service and/or CT scanner(s) other than an existing dental CT scanner service and/or dental CT scanner(s) or hospital-based portable CT scanner(s)~~

~~Sec. 9. (1) An applicant proposing to relocate an existing fixed CT scanner service shall demonstrate that the proposed project meets all of the following:~~

- ~~— (a) The existing fixed CT scanner service to be relocated has been in operation for at least 36 months as of the date an application is submitted to the Department.~~
- ~~— (b) The proposed new site is in the relocation zone.~~
- ~~— (c) The requirements of sections 5 or 7, as applicable, have been met.~~
- ~~— (d) The CT scanner service to be relocated performed at least an average of 7,500 CT-equivalents per fixed scanner in the most recent 12-month period for which the Department has verifiable data.~~
- ~~— (e) The applicant agrees to operate the CT scanner service in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.~~

~~— (2) An applicant proposing to relocate a fixed CT scanner(s) of an existing CT scanner service shall demonstrate that the proposed project meets all of the following:~~

- ~~— (a) The existing CT scanner service from which the CT scanner(s) is to be relocated has been in operation for at least 36 months as of the date an application is submitted to the Department.~~
- ~~— (b) The proposed new site is in the relocation zone.~~
- ~~— (c) The requirements of sections 5 or 7, as applicable, have been met.~~
- ~~— (d) Each existing CT scanner at the service from which a scanner is to be relocated performed at least an average of 7,500 CT-equivalents per fixed scanner in the most recent 12-month period for which the Department has verifiable data.~~
- ~~— (e) The applicant agrees to operate the CT scanner(s) at the proposed site in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.~~

~~Section 10. Requirements for approval for applicants proposing to relocate an existing dental CT scanner service and/or dental CT scanner(s)~~

~~Sec. 10. (1) An applicant proposing to relocate an existing fixed dental CT scanner service shall demonstrate that the proposed project meets all of the following:~~

- ~~— (a) The existing fixed dental CT scanner service to be relocated has been in operation for at least 36 months as of the date an application is submitted to the Department.~~
- ~~— (b) The proposed new site is in the relocation zone.~~
- ~~— (c) The requirements of sections 6 or 8, as applicable, have been met.~~
- ~~— (d) The dental CT scanner service to be relocated performed at least an average of 200 dental CT examinations per fixed dental CT scanner in the most recent 12-month period for which the Department has verifiable data.~~
- ~~— (e) The applicant agrees to operate the dental CT scanner service in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.~~

~~— (2) An applicant proposing to relocate a fixed dental CT scanner(s) of an existing dental CT scanner service shall demonstrate that the proposed project meets all of the following:~~

- ~~— (a) The existing dental CT scanner service from which the dental CT scanner(s) is to be relocated has been in operation for at least 36 months as of the date an application is submitted to the Department.~~
- ~~— (b) The proposed new site is in the relocation zone.~~
- ~~— (c) The requirements of sections 6 or 8, as applicable have been met.~~
- ~~— (d) Each existing dental CT scanner at the service from which a scanner is to be relocated performed at least an average of 200 dental CT examinations per fixed dental CT scanner in the most recent 12-month period for which the Department has verifiable data.~~
- ~~— (e) The applicant agrees to operate the dental CT scanner(s) at the proposed site in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.~~

Formatted: Tab stops: Not at 3.5" + 4.5" + 5" + 5.5" + 6" + 6.5" + 7" + 7.5"

Section 449. Requirements for approval for applicants proposing to acquire an existing CT scanner service or an existing CT scanner(s) other than an existing dental CT scanner service and/or an existing dental CT scanner(s) or hospital-based portable CT scanner(s)

Formatted: Line spacing: At least 12 pt, Don't hyphenate, Tab stops: 0.63", Left + Not at 0"

Sec. 449. (1) An applicant proposing to acquire an existing fixed or mobile CT scanner service, EXCEPT FOR AN APPLICANT APPROVED UNDER SECTION 3(1), shall demonstrate that a proposed project meets all of the following:

(a) The requirements of sections 5, 7, or 9, as applicable, have been met.
(b) For an application for the proposed first acquisition of an existing fixed or mobile CT scanner service, for which a final decision has not been issued after June 4, 2004, an existing CT scanner service to be acquired shall not be required to be in compliance with the volume requirement applicable to the seller/lessor on the date the acquisition occurs. The CT scanner service shall be operating at the applicable volume requirements set forth in Section 19 of these standards in the second 12 months after the date the service is acquired, and annually thereafter.

(c) For any application for proposed acquisition of an existing fixed or mobile CT scanner service, an applicant shall be required to demonstrate that the CT scanner service to be acquired performed at least 7,500 CT equivalents in the most recent 12-month period for which the Department has verifiable data.

(2) An applicant proposing to acquire an existing fixed or mobile CT scanner(s) of an existing fixed or mobile CT scanner service shall demonstrate that the proposed project meets all of the following:

(a) The requirements of sections 5, 7 or 9, as applicable, have been met.
(b) For any application for proposed acquisition of an existing fixed or mobile CT scanner(s) of an existing fixed or mobile CT scanner service, an applicant shall be required to demonstrate that the fixed or mobile CT scanner(s) to be acquired performed at least 7,500 CT equivalents in the most recent 12-month period for which the Department has verifiable data.

Section 4210. Requirements for approval for applicants proposing to acquire an existing dental CT scanner service or an existing dental CT scanner(s)

Sec. 4210. (1) An applicant proposing to acquire an existing fixed OR MOBILE dental CT scanner service shall demonstrate that a proposed project meets all of the following:

(a) The requirements of sections 6, 8, or 10, as applicable, have been met.
(b) For an application for the proposed first acquisition of an existing fixed OR MOBILE dental CT scanner service, for which a final decision has not been issued after the effective date of these standards, an existing dental CT scanner service to be acquired shall not be required to be in compliance with the volume requirement applicable to the seller/lessor on the date the acquisition occurs. The dental CT scanner service shall be operating at the applicable volume requirements set forth in Section 19 of these standards in the second 12 months after the date the service is acquired, and annually thereafter.

(c) For any application for proposed acquisition of an existing fixed OR MOBILE dental CT scanner service, an applicant shall be required to demonstrate that the CT scanner service to be acquired performed at least 200 dental CT examinations in the most recent 12-month period, for which the Department has verifiable data.

(2) An applicant proposing to acquire an existing fixed dental CT scanner(s) of an existing fixed OR MOBILE dental CT scanner service shall demonstrate that the proposed project meets all of the following:

(a) The requirements of sections 6, 8, or 10, as applicable, have been met.
(b) For any application for proposed acquisition of an existing fixed OR MOBILE dental CT scanner(s) of an existing fixed OR MOBILE dental CT scanner service, an applicant shall be required to demonstrate that the fixed OR MOBILE dental CT scanner(s) to be acquired performed at least 200 dental CT examinations in the most recent 12-month period for which the Department has verifiable data.

SECTION 11. REQUIREMENTS FOR A DEDICATED RESEARCH FIXED CT SCANNER

Formatted: Font: Not Bold, Highlight

Formatted: Font: Not Bold, Highlight

Formatted: Font: Not Bold, Highlight

484 SEC. 11. AN APPLICANT PROPOSING TO ADD A FIXED CT SCANNER TO AN EXISTING CT
485 SCANNER SERVICE FOR EXCLUSIVE RESEARCH USE SHALL DEMONSTRATE THE FOLLOWING:

Formatted: Font: Not Bold, Highlight

486
487 (1) THE APPLICANT AGREES THAT THE DEDICATED RESEARCH CT SCANNER WILL BE
488 USED PRIMARILY (70% OR MORE OF THE SCANS) FOR RESEARCH PURPOSES ONLY.

Formatted: Tab stops: 0.19", Left + 0.69",
Left + 0.75", Left + 0.88", Left + 1", Left +
Not at 0.94"

Formatted: Font: Not Bold, Highlight

489
490 (2) THE DEDICATED RESEARCH CT SCANNER SHALL OPERATE UNDER A PROTOCOL
491 APPROVED BY THE APPLICANT'S INSTITUTIONAL REVIEW BOARD, AS DEFINED BY PUBLIC LAW
492 93-348 AND REGULATED BY TITLE 45 CFR 46.

Formatted: Font: Not Bold, Highlight

Formatted: Font: Not Bold, Highlight

493
494 (3) THE PROPOSED SITE CAN HAVE NO MORE THAN THREE DEDICATED RESEARCH FIXED
495 CT SCANNERS APPROVED UNDER THIS SECTION.

Formatted: Font: Not Bold, Highlight

496
497 (4) THE DEDICATED RESEARCH SCANNER APPROVED UNDER THIS SECTION MAY NOT
498 UTILIZE CT PROCEDURES PERFORMED ON THE DEDICATED CT SCANNER TO DEMONSTRATE
499 NEED OR TO SATISFY CT CON REVIEW STANDARDS REQUIREMENTS.

Formatted: Highlight

Comment [A4]: Response to #3 CT
Workgroup Charge

Formatted: Font: Not Bold

500
501 Section 12. Requirements for approval of an applicant proposing a CT scanner used for the sole
502 purpose of performing dental CT examinations exclusively for research

Formatted: Tab stops: 0.19", Left + 0.69",
Left + 0.75", Left + 0.88", Left + 1", Left +
Not at 0.94"

503 Sec. 12. (1) An applicant proposing a CT scanner used for the sole purpose of performing dental CT
504 examinations exclusively for research shall demonstrate each of the following:

505 (a) The applicant operates a dental radiology program in a certified dental school.

506 (b) The research dental CT scanner shall operate under a protocol approved by the applicant's
507 institutional review board.

508 (c) The applicant agrees to operate the research dental CT scanner in accordance with the terms of
509 approval in Section 19(4).

510 (2) An applicant meeting the requirements of subsection (1) shall also demonstrate compliance with
511 the requirements of sections 4(2), 4(4) and 4(5).

512
513
514 Section 13. ~~Pilot program requirements~~ Requirements for approval of a hospital-based portable
515 CT scanner for initiation, expansion, replacement, and acquisition

Formatted: Highlight

516
517 Sec. 13. ~~As a pilot program, a~~ An applicant proposing to initiate, expand, replace, or acquire a hospital-
518 based portable CT scanner shall demonstrate that it meets all of the following:

Formatted: Highlight

519
520 (1) An applicant is limited to the initiation, expansion, replacement, or acquisition of no more than two
521 hospital-based portable CT scanners.

522
523 (2) The proposed site is a hospital licensed under Part 215 of the Code.

524
525 (3) The hospital has been certified as a level I or level II trauma facility by the American College of
526 Surgeons, or has performed >100 craniotomies in the most recent 12- month period verifiable by the
527 Department.

528
529 (4) The applicant agrees to operate the hospital-based portable CT scanner in accordance with all
530 applicable project delivery requirements set forth in Section 19 of these standards.

531
532 (5) The approved hospital-based portable CT scanner will not be subject to CT volume requirements.

533
534 (6) The applicant may not utilize CT procedures performed on a hospital-based portable CT scanner
535 to demonstrate need or to satisfy CT CON review standards requirements.

536

(7) The Commission may decide to have the requirements of the pilot program described in this section become a permanent part of the CT scanner services standards. If the Commission does not take action to make the pilot program a permanent part of the standards, the provisions of Section 13, as part of a pilot program, will expire on December 31, 2016 and be of no further force and effect after December 31, 2016. Any applicant seeking to be part of the pilot program described in this section must submit its application on or before December 1, 2013. These provisions shall not be applicable to any application which has not been submitted by December 1, 2013.

Section 14. Requirements for approval of a PET/CT hybrid for initiation, expansion, replacement, and acquisition

Sec. 14. An applicant proposing to initiate, expand, replace, or acquire a PET/CT hybrid shall demonstrate that it meets all of the following:

(1) There is an approved PET CON for the PET/CT hybrid, and the PET/CT hybrid is in compliance with all applicable project delivery requirements as set forth in the CON review standards for PET.

(2) The applicant agrees to operate the PET/CT hybrid in accordance with all applicable project delivery requirements set forth in Section 19 of these standards.

(3) The approved PET/CT hybrid will not be subject to CT volume requirements.

(4) A PET/CT scanner hybrid approved under the CON Review Standards for PET Scanner Services and the Review Standards for CT Scanner Services may not utilize CT procedures performed on a hybrid scanner to demonstrate need or to satisfy CT CON review standards requirements.

SECTION 15. REQUIREMENTS FOR APPROVAL OF A CT-ANGIO HYBRID UNIT FOR INITIATION, REPLACEMENT, AND ACQUISITION

SEC. 15. AN APPLICANT PROPOSING TO INITIATE, REPLACE, OR ACQUIRE A HOSPITAL-BASED CT-ANGIO HYBRID UNIT SHALL DEMONSTRATE EACH OF THE FOLLOWING, AS APPLICABLE TO THE PROPOSED PROJECT:

(1) THE PROPOSED SITE IS A LICENSED HOSPITAL UNDER PART 215 OF THE CODE.

(2) THE PROPOSED SITE HAS AN EXISTING FIXED CT SCANNER SERVICE THAT HAS BEEN OPERATIONAL FOR THE PREVIOUS 36 CONSECUTIVE MONTHS AND IS MEETING ITS MINIMUM VOLUME REQUIREMENTS.

(3) THE PROPOSED SITE OFFERS THE FOLLOWING SERVICES:

(a) DIAGNOSTIC CARDIAC CATHETERIZATION; OR

(b) INTERVENTIONAL RADIOLOGY; OR

(c) SURGICAL SERVICES

(4) THE PROPOSED CT-ANGIO HYBRID UNIT MUST BE LOCATED IN ONE OF THE FOLLOWING ROOMS:

(a) CARDIAC CATHETERIZATION LAB; OR

(b) INTERVENTIONAL RADIOLOGY SUITE; OR

(c) LICENSED OPERATING ROOM

(5) DIAGNOSTIC CT STUDIES SHALL NOT BE PERFORMED ON A CT-ANGIO HYBRID UNIT APPROVED UNDER THIS SECTION UNLESS THE PATIENT IS CURRENTLY UNDERGOING A CT-ANGIO HYBRID INTERVENTIONAL PROCEDURE AND IS IN NEED OF A SECONDARY DIAGNOSTIC CT STUDY.

(6) THE APPROVED CT-ANGIO HYBRID SHALL NOT BE SUBJECT TO CT VOLUME REQUIREMENTS.

(7) THE APPLICANT SHALL NOT UTILIZE THE PROCEDURES PERFORMED ON THE CT-ANGIO HYBRID UNIT TO DEMONSTRATE NEED OR TO SATISFY CT CON REVIEW STANDARDS REQUIREMENTS.

Comment [A5]: Response to Workgroup Charge #6 (which was added at the Workgroup meeting)

Section ~~45~~16. Additional requirements for approval of a mobile CT scanner service

Sec. ~~45~~16. (1) An applicant proposing to initiate a mobile CT scanner service in Michigan shall demonstrate that it meets all of the following:

(a) A separate CON application shall be submitted by the central service coordinator and each Michigan host facility.

(b) The normal route schedule, the procedures for handling emergency situations, and copies of all potential contracts related to the mobile CT scanner service shall be included in the CON application submitted by the central service coordinator.

(c) The requirements of sections 3, 5, or 7, as applicable, have been met.

(2) An applicant proposing to become a host facility on an existing mobile CT scanner network shall demonstrate that it meets all of the following:

(a) Approval of the application will not result in an increase in the number of operating mobile CT scanners for the mobile CT scanner network unless the requirements of Section 5 have been met.

(b) A separate CON application has been filed for each host facility.

(3) An applicant proposing to replace a central service coordinator on an existing mobile CT scanner network shall demonstrate that approval of the application will not replace the CT scanner and transporting equipment unless the applicable requirements of Section 7 have been met.

SECTION 17. ADDITIONAL REQUIREMENTS FOR APPROVAL OF A MOBILE DENTAL CT SCANNER SERVICE

SEC. 17. (1) AN APPLICANT PROPOSING TO INITIATE A MOBILE DENTAL CT SCANNER SERVICE IN MICHIGAN SHALL DEMONSTRATE THAT IT MEETS ALL OF THE FOLLOWING:

(A) A SEPARATE CON APPLICATION SHALL BE SUBMITTED BY THE CENTRAL SERVICE COORDINATOR AND EACH MICHIGAN HOST FACILITY.

(B) THE NORMAL ROUTE SCHEDULE, THE PROCEDURES FOR HANDLING EMERGENCY SITUATIONS, AND COPIES OF ALL POTENTIAL CONTRACTS RELATED TO THE MOBILE DENTAL CT SCANNER SERVICE SHALL BE INCLUDED IN THE CON APPLICATION SUBMITTED BY THE CENTRAL SERVICE COORDINATOR.

(C) THE REQUIREMENTS OF SECTIONS 4, 6 OR 8, AS APPLICABLE, HAVE BEEN MET.

(2) AN APPLICANT PROPOSING TO BECOME A HOST FACILITY ON AN EXISTING MOBILE DENTAL CT SCANNER NETWORK SHALL DEMONSTRATE THAT IT MEETS ALL OF THE FOLLOWING:

(A) APPROVAL OF THE APPLICATION WILL NOT RESULT IN AN INCREASE IN THE NUMBER OF OPERATING MOBILE DENTAL CT SCANNERS FOR THE MOBILE DENTAL CT SCANNER NETWORK UNLESS THE REQUIREMENTS OF SECTION 6 HAVE BEEN MET.

(B) A SEPARATE CON APPLICATION HAS BEEN FILED FOR EACH HOST FACILITY.

(3) AN APPLICANT PROPOSING TO REPLACE A CENTRAL SERVICE COORDINATOR ON AN EXISTING MOBILE DENTAL CT SCANNER NETWORK SHALL DEMONSTRATE THAT APPROVAL OF THE APPLICATION WILL NOT REPLACE THE DENTAL CT SCANNER AND TRANSPORTING EQUIPMENT UNLESS THE APPLICABLE REQUIREMENTS OF SECTION 8 HAVE BEEN MET.

~~Section 1615. Requirements for approval of an applicant proposing a CT scanner used for the sole purpose of performing dental CT examinations exclusively for research~~

~~Sec. 1615. (1) An applicant proposing a CT scanner used for the sole purpose of performing dental CT examinations exclusively for research shall demonstrate each of the following:~~

- ~~— (a) The applicant operates a dental radiology program in a certified dental school.~~
- ~~— (b) The research dental CT scanner shall operate under a protocol approved by the applicant's institutional review board.~~
- ~~— (c) The applicant agrees to operate the research dental CT scanner in accordance with the terms of approval in Section 19(4).~~
- ~~— (2) An applicant meeting the requirements of subsection (1) shall also demonstrate compliance with the requirements of sections 4(2), 4(4) and 4(5).~~

Section 4718. Requirements for approval of an applicant proposing to establish dedicated pediatric CT Scanner

Sec. 4718. (1) An applicant proposing to establish dedicated pediatric CT shall demonstrate all of the following:

- (a) The applicant shall have experienced at least 7,000 pediatric (< 18 years old) discharges (excluding normal newborns) in the most recent year of operation.
- (b) The applicant shall have performed at least 5,000 pediatric (< 18 years old) surgeries in the most recent year of operation.
- (c) The applicant shall have an active medical staff, at the time the application is submitted to the Department that includes, but is not limited to, physicians who are fellowship-trained in the following pediatric specialties:
 - (i) pediatric radiology (at least two)
 - (ii) pediatric anesthesiology
 - (iii) pediatric cardiology
 - (iv) pediatric critical care
 - (v) pediatric gastroenterology
 - (vi) pediatric hematology/oncology
 - (vii) pediatric neurology
 - (viii) pediatric neurosurgery
 - (ix) pediatric orthopedic surgery
 - (x) pediatric pathology
 - (xi) pediatric pulmonology
 - (xii) pediatric surgery
 - (xiii) neonatology
- (d) The applicant shall have in operation the following pediatric specialty programs at the time the application is submitted to the Department:
 - (i) pediatric bone marrow transplant program
 - (ii) established pediatric sedation program
 - (iii) pediatric open heart program

(2) An applicant meeting the requirements of subsection (1) shall be exempt from meeting the requirements of Section 3 of these standards.

Section 4819. Requirements for MEDICAID approval ~~all applicants~~ PARTICIPATION

697 | Sec. ~~48~~19. An applicant shall provide verification of Medicaid participation. An applicant that is a new
698 | provider not currently enrolled in Medicaid shall certify that proof of Medicaid participation will be provided
699 | to the Department within six (6) months from the offering of services if a CON is approved.

701 | **Section 4920. Project delivery requirements --AND terms of approval for all applicants**

703 | ~~Sec. 4920. (1)~~ An applicant shall agree that, if approved, the ~~services provided by the~~ CT scanner(s)
704 | SERVICES shall be delivered in compliance with the following terms of ~~CON~~ approval:

Comment [A6]: CT Workgroup needs to Review the Project Delivery Requirements.

705 | (a1) Compliance with these standards.

706 | ~~(b) Compliance with applicable safety and operating standards~~

708 | (e2) Compliance with the following quality assurance standards:

709 | ~~(i) The approved CT scanners shall be operating at the applicable required volumes within the time~~
710 | ~~periods specified in these standards, and annually thereafter.~~

711 | (ia) The applicant shall establish a mechanism to assure that the CT scanner facility is staffed so that:

712 | (Ai) The screening of requests for CT procedures and interpretation of CT procedures will be
713 | performed by physicians with training and experience in the appropriate diagnostic use and interpretation
714 | of cross-sectional images of the anatomical region(s) to be examined, and

715 | (Eii) The CT scanner is operated by physicians and/or is operated by radiological technologists
716 | qualified by training and experience to operate the CT scanner safely and effectively.

717 | For purposes of evaluating (ia)(Ai), the Department shall consider it prima facie evidence of a
718 | satisfactory assurance mechanism as to screening and interpretation if the applicant requires the
719 | screening of requests for and interpretations of CT procedures to be performed by physicians who are
720 | board certified or eligible in radiology or are neurologists or other specialists trained in cross-sectional
721 | imaging of a specific organ system. For purposes of evaluating (ia)(Ei) the Department shall consider it
722 | prima facie evidence of a satisfactory assurance mechanism as to the operation of a CT scanner if the
723 | applicant requires the CT scanner to be operated by a physician or by a technologist registered by the
724 | American Registry of Radiological Technologists (ARRT) or the American Registry of Clinical Radiography
725 | Technologists (ARCRT). However, the applicant may submit and the Department may accept other
726 | evidence that the applicant has established a mechanism to assure that the CT scanner facility is
727 | appropriately and adequately staffed as to screening, interpretation, and/or operation of a CT scanner.

728 | (ib) The applicant shall employ or contract with a radiation physicist to review the quality and safety of
729 | the operation of the CT scanner.

730 | (ic) The applicant shall assure that at least one of the physicians responsible for the screening and
731 | interpretation as defined in subsection (ia)(Ai) will be in the CT facility or available on a 24-hour basis
732 | (either on-site or through telecommunication capabilities) to make the final interpretation.

733 | (vd) In the case of an urgent or emergency CT scan, the applicant shall assure that a physician so
734 | authorized by the applicant to interpret initial scans will be on-site or available through telecommunication
735 | capabilities within 1 hour following completion of the scanning procedure to render an initial interpretation
736 | of the scan. A final interpretation shall be rendered by a physician so authorized under subsection (ia)(Ai)
737 | within 24 hours.

738 | (ie) The applicant shall have, within the CT scanner facility, equipment and supplies to handle clinical
739 | emergencies that might occur within the CT unit, with CT facility staff trained in CPR and other appropriate
740 | emergency interventions, and a physician on site in or immediately available to the CT scanner at all times
741 | when patients are undergoing scans.

742 | (vif) Fixed CT scanner services at each facility shall be made available 24 hours a day for emergency
743 | patients.

744 | (viig) The applicant shall accept referrals for CT scanner services from all appropriately licensed
745 | practitioners.

746 | (xh) The applicant shall establish and maintain: (a) a standing medical staff and governing body (or its
747 | equivalent) requirement that provides for the medical and administrative control of the ordering and
748 | utilization of CT patient procedures, and (b) a formal program of utilization review and quality assurance.
749 | These responsibilities may be assigned to an existing body of the applicant, as appropriate.

(*) An applicant approved under Section 17 must be able to prove that all radiologists, technologists and nursing staff working with CT patients have continuing education or in-service training on pediatric low-dose CT. The site must also be able to provide evidence of defined low-dose pediatric CT protocols.

(3) COMPLIANCE WITH THE FOLLOWING ACCESS TO CARE REQUIREMENTS:

(*)a The applicant, to assure that the CT scanner will be utilized by all segments of the Michigan population, shall:

(A) not deny ANY CT scanner services to any individual based on ability to pay or source of payment;

(B) provide ALL CT scanning services to any individual based on the clinical indications of need for the service; and

(C) maintain information by payor and non-paying sources to indicate the volume of care from each source provided annually.

(b) An applicant shall participate in Medicaid at least 12 consecutive months within the first two years of operation and continue to participate annually thereafter.

(c) The operation of and referral of patients to the CT scanner shall be in conformance with 1978 PA 368, Sec. 16221, as amended by 1986 PA 319; MCL 333.16221; MSA 14.15 (16221).

Compliance with selective contracting requirements shall not be construed as a violation of this term.

(4) COMPLIANCE WITH THE FOLLOWING MONITORING AND REPORTING REQUIREMENTS:

(ja) The approved CT scanners shall be operating at the applicable required volumes AN AVERAGE within the time periods specified in these standards, OF 7,500 CT EQUIVALENTS SCANNER PER FIXED SCANNER AND 3,500 CT EQUIVALENTS PER MOBILE SCANNER PER YEAR FOR THE SECOND 12-MONTH PERIOD AFTER BEGINNING OPERATION OF THE CT SCANNER, and annually thereafter.

Formatted: Strikethrough

(*)b The applicant shall participate in a data collection network established and administered by the Department or its designee. The data may include, but is not limited to, annual budget and cost information, operating schedules, through-put schedules, demographic and diagnostic information, the volume of care provided to patients from all payor sources, and other data requested by the Department, and approved by the Commission. The applicant shall provide the required data on a separate basis for each separate and distinct site as required by the Department; in a format established by the Department; and in a mutually agreed upon media. The Department may elect to verify the data through on-site review of appropriate records.

(*)c Equipment to be replaced shall be removed from service.

(*)d The applicant shall provide the Department with a TIMELY notice ~~stating the date the approved CT scanner service is placed in operation and such notice shall be submitted to the Department OF THE PROPOSED PROJECT IMPLEMENTATION~~ consistent with applicable statute and promulgated rules.

~~(xv) An applicant shall participate in Medicaid at least 12 consecutive months within the first two years of operation and continue to participate annually thereafter.~~

(e) An applicant approved under Section 4 shall not be required to be in compliance with subsection

(e2), ~~but shall be in compliance with the following quality assurance standards:~~

~~(i) The CT scanner shall be operating at least 200 CT equivalents per year for the second 12-month period after beginning operation of the dental CT scanner and annually thereafter.~~

(5) COMPLIANCE WITH THE FOLLOWING DENTAL CT SCANNER REQUIREMENTS, IF APPLICABLE:

(*)a The CT scanner will be used for the sole purpose of dental CT examinations.

(*)b The applicant shall demonstrate to the satisfaction of the Department that the person(s) (e.g., technician, dentist) operating the dental CT scanner has been appropriately trained and/or certified by one of the following groups, as recognized by the Department: a dental radiology program in a certified dental school, an appropriate professional society, or a dental continuing education program accredited by the American Dental Association.

(*)c The applicant shall demonstrate to the satisfaction of the Department that the dental CT examinations generated by the dental CT scanner will be interpreted by a licensed dentist(s) trained

and/or certified by one of the following groups, as recognized by the Department: a dental radiology program in a certified dental school, an appropriate professional society, or a dental continuing education program accredited by the American Dental Association.

(v)d) The applicant shall demonstrate to the satisfaction of the Department that the dentists using the dental CT examinations for performing dental procedures has had the appropriate training and/or experience certified by one of the following groups, as recognized by the Department: a dental radiology program in a certified dental school, an appropriate professional society, or a dental continuing education program accredited by the American Dental Association.

(v)e) The applicant, to assure that the dental CT scanner will be utilized by all segments of the Michigan population, shall:

(A)i) not deny dental CT scanner services to any individual based on ability to pay or source of payment;

(B)ii) provide dental CT scanning services to any individual based on the clinical indications of need for the service; and

(C)iii) maintain information by payor and non-paying sources to indicate the volume of care from each source provided annually. Compliance with selective contracting requirements shall not be construed as a violation of this term.

(f) The CT scanner shall be operating at least 200 CT equivalents per year for the second 12-month period after beginning operation of the dental CT scanner and annually thereafter.

(vii)g) The applicant shall participate in a data collection network established and administered by the Department or its designee. The data may include, but is not limited to, annual budget and cost information, operating schedules, through-put schedules, demographic and diagnostic information, the volume of care provided to patients from all payor sources, and other data requested by the Department, and approved by the Commission. The applicant shall provide the required data on a separate basis for each separate and distinct site as required by the Department; in a format established by the Department; and in a mutually agreed upon media. The Department may elect to verify the data through on-site review of appropriate records.

(viii)h) Equipment to be replaced shall be removed from service.

(ix)i) The applicant shall provide the Department with a TIMELY notice ~~stating the date the approved dental CT scanner service is placed in operation and such notice shall be submitted to the Department~~ OF THE PROPOSED PROJECT IMPLEMENTATION consistent with applicable statute and promulgated rules.

(x)j) An applicant shall participate in Medicaid at least 12 consecutive months within the first two years of operation and continue to participate annually thereafter.

~~(2e)i) The agreements and assurances required by this section shall be in the form of a certification agreed to by the applicant or its authorized agent.~~

~~— (3) — The operation of and referral of patients to the CT scanner shall be in conformance with 1978 PA 368, Sec. 16221, as amended by 1986 PA 319; MCL 333.16221; MSA 14.15 (16221).~~

(46) An applicant for a CT scanner used for dental research under Section ~~46~~18(1) shall agree that the services provided by the CT scanner approved pursuant to Section 16(1) shall be delivered in compliance with the following terms of CON approval:

(a) The capital and operating costs relating to the CT scanner used for dental research pursuant to Section ~~46~~12(1) shall be charged only to a specific research account(s) and not to any patient or third-party payor.

(b) The CT scanner used for dental research approved pursuant to Section ~~46~~12(1) shall not be used for any purposes other than as approved by the institutional review board unless the applicant has obtained CON approval for the CT scanner pursuant to part 222 and these standards, other than Section ~~46~~12.

(57) An applicant approved under Section 13 shall be in compliance with the following:

(a) Portable CT scanner can only be used by a qualifying pilot program for the following purposes:

(i) Brain scanning of patients being treated in an adult or pediatric Intensive Care Unit (ICU).

Formatted: Not Highlight

- 858 (ii) Non-diagnostic, intraoperative guidance in an operating room.
859 (b) The approved applicant must provide annual reports to the Department by January 31st of each
860 year for the preceding calendar year. This requirement applies to all applicants approved under Section
861 13 and begins with 2010 data which is to be reported in 2011.
862 (c) The following data must be reported to the Department:
863 (i) Number of adult studies (age>=18)
864 (ii) Number of pediatric studies (age<18)
865 (iii) Number of studies performed using a portable CT on the same patient while that patient is in an
866 ICU

867 ~~(iv) Number of patients scanned on a portable CT that underwent subsequent scanning on a fixed CT~~
868 ~~within 12 hours of the portable CT scan~~

Formatted: Highlight

869
870 (8) AN APPLICANT APPROVED UNDER SECTION 15 SHALL BE IN COMPLIANCE WITH THE
871 FOLLOWING:

872 (a) THE PROPOSED SITE OFFERS THE FOLLOWING SERVICES:

873 (i) DIAGNOSTIC CARDIAC CATHETERIZATION; OR

874 (ii) INTERVENTIONAL RADIOLOGY; OR

875 (iii) SURGICAL SERVICES

876 (b) THE PROPOSED CT-ANGIO HYBRID UNIT MUST BE LOCATED IN ONE OF THE
877 FOLLOWING ROOMS:

878 (i) CARDIAC CATHETERIZATION LAB; OR

879 (ii) INTERVENTIONAL RADIOLOGY SUITE; OR

880 (iii) LICENSED OPERATING ROOM

881
882 (29) The agreements and assurances required by this section shall be in the form of a certification
883 agreed to by the applicant or its authorized agent.

884
885 **Section 2021. Project delivery requirements AND additional terms of approval for applicants**
886 **involving mobile CT scanners**

887
888 | Sec. 2021. (1) In addition to the provisions of Section 19, an applicant for a mobile CT scanner shall
889 agree that the services provided by the mobile CT scanner(s) shall be delivered in compliance with the
890 following terms of CON approval:

891 (a) A host facility shall submit only one CON application for a CT scanner for review at any given
892 time.

893 (b) A mobile CT scanner with an approved CON shall notify the Michigan Department of Community
894 Health prior to ending service with an existing host facility.

895 (c) A CON shall be required to add a host facility.

896 (d) A CON shall be required to change the central service coordinator.

897 (e) Each host facility must have at least one board certified or board eligible radiologist on its medical
898 staff. The radiologist(s) shall be responsible for: (i) establishing patient examination and infusion
899 protocol, and (ii) providing for the interpretation of scans performed by the mobile CT scanner.

900 (f) Each mobile CT scanner service must have an Operations Committee with members
901 representing each host facility, the central service coordinator, and the central service medical director.
902 This committee shall oversee the effective and efficient use of the CT scanner, establish the normal route
903 schedule, identify the process by which changes are to be made to the schedule, develop procedures for
904 handling emergency situations, and review the ongoing operations of the mobile CT scanner on at least a
905 quarterly basis.

906 (g) The central service coordinator shall arrange for emergency repair services to be available 24
907 hours each day for the mobile CT scanner as well as the vehicle transporting the equipment. In addition,
908 to preserve image quality and minimize CT scanner downtime, calibration checks shall be performed on
909 the CT scanner at least once each work day and routine maintenance services shall be provided on a
910 regularly scheduled basis, at least once a week during hours not normally used for patient procedures.

(h) Each host facility must provide a properly prepared parking pad for the mobile CT scanner of sufficient load-bearing capacity to support the vehicle, a waiting area for patients, and a means for patients to enter the vehicle without going outside (such as a canopy or enclosed corridor). Each host facility must also provide the capability for processing the film and maintaining the confidentiality of patient records. A communication system must be provided between the mobile vehicle and each host facility to provide for immediate notification of emergency medical situations.

(i) A mobile CT scanner service shall operate under a contractual agreement that includes the provision of CT scanner services at each host facility on a regularly scheduled basis.

(j) The volume of utilization at each host facility shall be reported to the Department by the central service coordinator under the terms of Section 19(1)(c)(xi).

(2) The agreements and assurances required by this section shall be in the form of a certification agreed to by the applicant or its authorized agent.

Section 24.22. Determination of CT Equivalents

Sec. 24.22. CT equivalents shall be calculated as follows:

(a) Each billable procedure for the time period specified in the applicable section(s) of these standards shall be assigned to a category set forth in Table 1.

(b) The number of billable procedures for each category in the time period specified in the applicable section(s) of these standards shall be multiplied by the corresponding conversion factor in Table 1 to determine the number of CT equivalents for that category for that time period.

(c) The number of CT equivalents for each category shall be summed to determine the total CT equivalents for the time period specified in the applicable section(s) of these standards.

(d) The conversion factor for pediatric/special needs patients does not apply to procedures performed on a dedicated pediatric CT scanner.

Category	Number of Billable CT Procedures		Conversion Factor		CT Equivalents
Head Scans w/o Contrast (includes dental CT examinations)	_____	X	1.00	=	_____
Head Scans with Contrast	_____	X	1.25	=	_____
Head Scans w/o & w Contrast	_____	X	1.75	=	_____
Body Scans w/o Contrast	_____	X	1.50	=	_____
Body Scans with Contrast	_____	X	1.75	=	_____
Body Scans w/o & w Contrast	_____	X	2.75	=	_____
Pediatric/Special Needs Patient Head scans w/o Contrast (includes dental CT examinations)	_____	x	1.25	=	_____
Pediatric/Special Needs Patient Head Scans with Contrast	_____	x	1.50	=	_____
Pediatric/Special Needs Patient Head Scans w/o & with Contrast	_____	x	2.00	=	_____
Pediatric/Special Needs Patient Body Scans w/o Contrast	_____	x	1.75	=	_____
Pediatric/Special Needs Patient Body Scans with Contrast	_____	x	2.00	=	_____
Pediatric/Special Needs Patient Body Scans w/o & with Contrast	_____	x	3.00	=	_____
Total CT Equivalents					_____

965
966 | **Section 2223. Documentation of projections**
967
968 | Sec. 2223. An applicant required to project volumes under sections 3, AND 4 and 5 shall demonstrate
969 the following, as applicable:
970 (1) An applicant required to project under Section 3 shall demonstrate that the projection is based on
971 historical physician referrals that resulted in an actual scan for the most recent 12-month period
972 immediately preceding the date of the application. Historical physician referrals will be verified with the
973 data maintained by the Department through its "Annual Hospital statistical survey" and/or "Annual
974 Freestanding Statistical Survey."
975
976 (2) An applicant required to project under Section 4 shall demonstrate that the projection is based on
977 a combination of the following for the most recent 12-month period immediately preceding the date of the
978 application:
979 (a) the number of dental procedures performed by the applicant, and
980 (b) the number of committed dental procedures performed by referring licensed dentists. Further, the
981 applicant and the referring licensed dentists shall substantiate the numbers through the submission of
982 HIPAA compliant billing records.
983
984 (3) An applicant required to project under Section 5 shall demonstrate that the projection is based on
985 historical utilization at the applicant's site for the most recent 12-month period immediately preceding the
986 date of the application.
987
988 (4) An applicant shall demonstrate that the projected number of referrals to be performed at the
989 | proposed site under subsections (1) ~~and (2)~~ are from an existing CT scanner service that is in compliance
990 with the volume requirements applicable to that service, and will continue to be in compliance with the
991 volume requirements applicable to that service subsequent to the initiation of the proposed CT scanner
992 | service by an applicant. THIS DOES NOT INCLUDE DENTAL CT SCANNERS. Only excess CT
993 equivalents equal to or greater than what is being committed pursuant to this subsection may be used to
994 document projections under subsection (1). In demonstrating compliance with this subsection, an
995 applicant shall provide each of the following:
996 (a) A written commitment from each referring physician that he or she will refer at least the volume of
997 CT scans to be transferred to the proposed CT scanner service for no less than 3 years subsequent to the
998 initiation of the CT scanner service proposed by an applicant.
999 (b) The number of referrals committed must have resulted in an actual CT scan of the patient at the
1000 existing CT scanner service from which referral will be transferred. The committing physician must make
1001 available HIPAA compliant audit material if needed upon Department request to verify referral sources and
1002 outcomes. Commitments must be verified by the most recent data set maintained by the Department
1003 through its "Annual Hospital Statistical Survey" and/or "Annual Freestanding Statistical Survey."
1004 (c) The projected referrals are from an existing CT scanner service within a 75-mile radius for rural
1005 and micropolitan statistical area counties or 20-mile radius for metropolitan statistical area counties.
1006

1007 | **Section 2324. Effect on prior CON review standards; comparative reviews**
1008

1009 | Sec. 2324. (1) These CON review standards supersede and replace the CON Review Standards
1010 for Computed Tomography Scanner Services approved by the CON Commission on ~~April 30,~~
1011 ~~2008~~ DECEMBER 15, 2011 and effective on ~~June 20, 2008~~ FEBRUARY 27, 2012.
1012
1013 | (2) Projects reviewed under these standards shall not be subject to comparative review.
1014

Formatted: Line spacing: single, Hyphenate,
Tab stops: Not at -0.75" + -0.5" + -0.17" +
0.15" + 0.52" + 0.96" + 2.13" + 2.5" +
3.33" + 4.02" + 4.83" + 5" + 5.31" + 6.02"
+ 6.4" + 7" + 7.5"

Formatted: Indent: First line: 0"

Section 24. Health Service Areas

Sec. 24. Counties assigned to each of the health service areas are as follows:

HEALTH SERVICE AREA	COUNTIES			
1 –Southeast	Livingston Macomb Wayne	Monroe Oakland	St. Clair Washtenaw	
2 –Mid-Southern	Clinton Eaton	Hillsdale Ingham	Jackson Lenawee	
3 –Southwest	Barry Berrien Branch	Calhoun Cass Kalamazoo	St. Joseph Van Buren	
4 –West	Allegan Ionia Kent Lake	Mason Mecosta Montcalm Muskegon	Newaygo Oceana Osceola Ottawa	
5 –GLS	Genesee	Lapeer	Shiawassee	
6 –East	Arenac Bay Clare Gladwin Gratiot	Huron Iosco Isabella Midland Ogemaw	Roscommon Saginaw Sanilac Tuscola	
7 –Northern-Lower	Alcona Alpena Antrim Benzie Charlevoix Cheboygan	Crawford Emmet Gd Traverse Kalkaska Leelanau Manistee	Missaukee Montmorency Oscoda Otsego Presque Isle Wexford	
8 –Upper Peninsula	Alger Baraga Chippewa Delta Dickinson	Gogebic Houghton Iron Keweenaw Luce	Mackinac Marquette Menominee Ontonagon Schoolcraft	

1059 |

APPENDIX A-B

1060

1061 Rural Michigan counties are as follows:

1062

1063	Alcona	Hillsdale	Ogemaw
1064	Alger	Huron	Ontonagon
1065	Antrim	Iosco	Osceola
1066	Arenac	Iron	Oscoda
1067	Baraga	Lake	Otsego
1068	Charlevoix	Luce	Presque Isle
1069	Cheboygan	Mackinac	Roscommon
1070	Clare	Manistee	Sanilac
1071	Crawford	Mason	Schoolcraft
1072	Emmet	Montcalm	Tuscola
1073	Gladwin	Montmorency	
1074	Gogebic	Oceana	

1075

1076 Micropolitan statistical area Michigan counties are as follows:

1077

1078	Allegan	Gratiot	Mecosta
1079	Alpena	Houghton	Menominee
1080	Benzie	Isabella	Midland
1081	Branch	Kalkaska	Missaukee
1082	Chippewa	Keweenaw	St. Joseph
1083	Delta	Leelanau	Shiawassee
1084	Dickinson	Lenawee	Wexford
1085	Grand Traverse	Marquette	

1086

1087 Metropolitan statistical area Michigan counties are as follows:

1088

1089	Barry	Ionia	Newaygo
1090	Bay	Jackson	Oakland
1091	Berrien	Kalamazoo	Ottawa
1092	Calhoun	Kent	Saginaw
1093	Cass	Lapeer	St. Clair
1094	Clinton	Livingston	Van Buren
1095	Eaton	Macomb	Washtenaw
1096	Genesee	Monroe	Wayne
1097	Ingham	Muskegon	

1098

1099 Source:

1100 65 F.R., p. 82238 (December 27, 2000)

1101 Statistical Policy Office

1102 Office of Information and Regulatory Affairs

1103 United States Office of Management and Budget