

CERTIFICATE OF NEED

ADVISORY

Michigan Department of Health & Human Services

CERTIFICATE OF NEED

South Grand Building
333 S. Grand Avenue
Lansing, Michigan 48933
(517) 241-3344 -Fax (517) 241-2962

Lithotripsy Shielding

A lithotripsy unit is to be operated in a room that complies with Radiation Safety Standards. Operating or procedure rooms in which a lithotripsy unit is used must have a steel or leaded door, as well as shielding for windows in these rooms. If a unit is to be used in an area that does not meet these criteria, approval of the radiation shielding must be obtained from the Radiation Safety Section prior to clinical use in that location. For detailed information, click on this link to the [Radiation Safety Web Site](#).