



600 South Wagner Road  
Ann Arbor, MI 48103-6019  
Phone: 734-913-6130  
Fax: 734-913-6103

### Sample Analysis Report

February, 2006

Analyst Initials NO  
Date 03/03/06

Sample Name - Date Sampled - Time Sampled	1,4-Dioxane Results (ppb)	R.L. (ppb)	Bromate Results (ppb)	R.L. (ppb)	Bromide Results (ppb)	R.L. (ppb)	Comments
<b>EXTRACTION WELLS</b>							
AE-3-02-06-06-08:36	79	1.0		5.0		10.0	
HZ-S-02-06-06-11:37	1508	1.0		5.0		10.0	
LB-1-02-06-06-08:32	418	1.0		5.0		10.0	
LB-2-02-06-06-08:34	968	1.0		5.0		10.0	
PW-1-02-06-06-07:44	1348	1.0		5.0		10.0	
TW-1-02-06-06-11:30	266	1.0		5.0		10.0	
DOLPH-02-06-06-11:01	112	1.0		5.0		10.0	
TW-3-02-06-06-11:32	149	1.0		5.0		10.0	
TW-5-02-06-06-11:18	1030	1.0		5.0		10.0	
TW-6-02-06-06-07:48	747	1.0		5.0		10.0	
TW-8-02-06-06-14:14	887	1.0		5.0		10.0	
TW-9-02-06-06-11:34	1905	1.0		5.0		10.0	
TW-10-02-06-06-11:27	3501	1.0		5.0		10.0	
TW-11-02-06-06-11:20	456	1.0		5.0		10.0	
TW-12-02-06-06-11:03	69	1.0		5.0		10.0	
TW-13-02-06-06-11:49	434	1.0		5.0		10.0	
TW-14-02-06-06-11:41	221	1.0		5.0		10.0	
TW-17-02-06-06-11:45	204	1.0		5.0		10.0	
<b>TEMPORARY EXTRACTION WELLS</b>							
MW-50-02-06-06-14:16	303	1.0		5.0		10.0	
SW-COMB-02-06-06-07:46	697	1.0		5.0		10.0	
<b>RED POND</b>							

Sample Name - Date Sampled - Time Sampled	1,4-Dioxane Results (ppb)	R.L. (ppb)	Bromate Results (ppb)	R.L. (ppb)	Bromide Results (ppb)	R.L. (ppb)	Comments
<b>RED POND</b>							
Red Pond-02-06-06-07:50	835	1.0		5.0	144	10.0	
Red Pond-02-13-06-07:28	1047	1.0		5.0	100	10.0	
Red Pond-02-20-06-07:25	1054	1.0		5.0	107	10.0	
Red Pond-02-27-06-07:25	1223	1.0		5.0	99	10.0	
<b>OUTFALL001</b>							
OUTFALL-02-01-06-	5	1.0	8	5.0		10.0	
OUTFALL-02-02-06-	4	1.0	8	5.0		10.0	
OUTFALL-02-05-06-	5	1.0	7	5.0		10.0	
OUTFALL-02-06-06-	4	1.0	6	5.0		10.0	
OUTFALL-02-07-06-	5	1.0	6	5.0		10.0	
OUTFALL-02-08-06-	4	1.0	6	5.0		10.0	
OUTFALL-02-09-06-	4	1.0	8	5.0		10.0	
OUTFALL-02-12-06-	5	1.0	8	5.0		10.0	
OUTFALL-02-13-06-	5	1.0	7	5.0		10.0	
OUTFALL-02-14-06-	5	1.0	6	5.0		10.0	
OUTFALL-02-15-06-	6	1.0	nd	5.0		10.0	
OUTFALL-02-16-06-	6	1.0	9	5.0		10.0	
OUTFALL-02-19-06-	5	1.0	9	5.0		10.0	
OUTFALL-02-20-06-	5	1.0	nd	5.0		10.0	
OUTFALL-02-21-06-	5	1.0	9	5.0		10.0	
OUTFALL-02-22-06-	5	1.0	8	5.0		10.0	
OUTFALL-02-23-06-	5	1.0	7	5.0		10.0	
OUTFALL-02-26-06-	5	1.0	7	5.0		10.0	
OUTFALL-02-27-06-	5	1.0	8	5.0		10.0	
OUTFALL-02-28-06-	5	1.0	9	5.0		10.0	
<b>C3</b>							
MW-5d-02-02-06-12:27	8304	1.0		5.0		10.0	
MW-11i-02-10-06-13:20	47	1.0		5.0		10.0	

Sample Name - Date Sampled - Time Sampled	1,4-Dioxane Results (ppb)	R.L. (ppb)	Bromate Results (ppb)	R.L. (ppb)	Bromide Results (ppb)	R.L. (ppb)	Comments
<b>C3</b>							
MW-37-02-02-06-14:22	203	1.0		5.0		10.0	
<b>D0</b>							
MW-53i-02-08-06-14:42	45	1.0		5.0		10.0	
A2 Cleaning Supply-02-14-06-15:37	113	1.0		5.0		10.0	
<b>D2</b>							
AE-2-02-07-06-15:00	34	1.0		5.0		10.0	
MW-38d-02-02-06-15:19	241	1.0		5.0		10.0	
MW-47s-02-09-06-11:50	nd	1.0		5.0		10.0	
MW-47d-02-09-06-12:15	nd	1.0		5.0		10.0	
MW-54d-02-08-06-13:45	11	1.0		5.0		10.0	
MW-56s-02-14-06-15:56	154	1.0		5.0		10.0	
MW-77-02-01-06-14:29	307	1.0		5.0		10.0	
545 Allison-02-08-06-11:33	4	1.0		5.0		10.0	
430 Barber West-02-01-06-11:38	105	1.0		5.0		10.0	
MW-400 Clarendon-02-01-06-10:46	nd	1.0		5.0		10.0	
<b>E</b>							
MW-72s-02-06-06-15:16	74	1.0		5.0		10.0	
MW-72d-02-06-06-15:53	3231	1.0		5.0		10.0	
MW-85-02-06-06-11:19	1699	1.0		5.0		10.0	
MW-87s-02-07-06-10:25	612	1.0		5.0		10.0	
MW-87d-02-07-06-10:06	697	1.0		5.0		10.0	
MW-88-02-07-06-11:08	1197	1.0		5.0		10.0	
MW-96-02-10-06-		1.0		5.0		10.0	
TW-15-02-08-06-11:09	97	1.0		5.0		10.0	
TW-18-02-06-06-11:12	1848	1.0		5.0		10.0	
<b>SW</b>							
MW-45s-02-02-06-13:17	144	1.0		5.0		10.0	
MW-78-02-02-06-11:37	40	1.0		5.0		10.0	

Sample Name - Date Sampled - Time Sampled	1,4-Dioxane Results (ppb)	R.L. (ppb)	Bromate Results (ppb)	R.L. (ppb)	Bromide Results (ppb)	R.L. (ppb)	Comments
<b>Not Determined</b>							
2601 Dexter-02-01-06-13:42	2	1.0		5.0		10.0	

nd=Not detected at or above the Reporting Limit (R.L.)

# 1,4-Dioxane Precision and Accuracy Control Charting

Analysis Date	Method Blank	CVS True Value	CVS Result	CVS % Recovery	LFB True Value:	LFB Result:	LFB % Recovery:	Sample Result:	MS/MSD True Value:	MS Result:	MSD Result:	MS % Recovery:	MSD % Recovery:	MS/MSD Mean:	MS/MSD RSD:	MS/MSD RPD:
2/1/2006	0	10.00	9.90	99	10.00	9.93	99.30	4.65	14.83	14.91	14.83	102.60	101.80	14.87	0.38	0.54%
					10.00	9.20	92.00									
2/2/2006	0	10.00	9.55	95.5	10.00	9.01	90.10	4.82								
2/3/2006	0	10.00	9.92	99.2	10.00	9.22	92.20									
	0	10.00	9.37	93.7	10.00	9.49	94.90	4.34	13.05	13.95	13.05	96.10	87.10	13.50	4.71	6.67%
2/6/2006	0	10.00	9.33	93.3	10.00	8.97	89.70									
2/7/2006	0	10.00	9.39	93.9	10.00	9.94	99.40	4.32	14.60	14.39	14.60	100.70	102.80	14.50	1.02	1.45%
	0	10.00	9.98	99.8	10.00	10.09	100.90									
2/8/2006	0	10.00	9.57	95.7	10.00	9.72	97.20	4.64	14.58	14.25	14.58	96.10	99.40	14.42	1.62	2.29%
	0	10.00	9.99	99.9	10.00	9.86	98.60									
2/9/2006	0	10.00	9.72	97.2	10.00	10.12	101.20									
	0	10.00	9.62	96.2	10.00	9.56	95.60	4.59	14.22	14.25	14.22	96.60	96.30	14.24	0.15	0.21%
2/10/2006	0	10.00	9.55	95.5	10.00	9.08	90.80									
2/13/2006	0	10.00	9.37	93.7	10.00	9.38	93.80									
2/14/2006	0	10.00	9.38	93.8	10.00	9.20	92.00									
2/15/2006	0	10.00	9.64	96.4	10.00	9.25	92.50	4.83	14.71	14.59	14.71	97.60	98.80	14.65	0.58	0.82%
2/16/2006	0	10.00	9.55	95.5	10.00	9.28	92.80	5.60	15.28	15.27	15.28	96.70	96.80	15.28	0.05	0.07%
2/20/2006	0	10.00	9.60	96	10.00	9.33	93.30	5.43	15.33	15.28	15.33	98.50	99.00	15.31	0.23	0.33%
2/21/2006	0	10.00	9.34	93.4	10.00	9.32	93.20	5.38	13.39	14.29	13.39	89.10	80.10	13.84	4.60	6.50%
2/22/2006	0	10.00	8.83	88.3	10.00	8.61	86.10	5.45	15.68	14.28	15.68	88.30	102.30	14.98	6.61	9.35%
2/24/2006	0	10.00	9.04	90.4	10.00	8.98	89.80									
2/27/2006	0	10.00	9.04	90.4	10.00	9.03	90.30	5.33	15.03	14.96	15.03	96.30	97.00	15.00	0.33	0.47%
2/28/2006	0	10.00	8.54	85.4	10.00	8.62	86.20	5.36	14.45	14.58	14.45	92.20	90.90	14.52	0.63	0.90%

CVS Mean: 9.46    CVS Standard Dev: 0.36    3 Standard Dev.: 1.09    Lower Control Limit: 8.37    Upper Control Limit: 10.56